



Unit D1 Clane Business Park,
College Road. Clane,
Co. Kildare.
Ph: +353 1 835 3084

WEIGHBRIDGE CALIBRATION TEST REPORT NO : 22955-C1

CUSTOMER:	ASH Wrexham Recycling Centre	TYPE APPROVAL CERT NO:	#REF!
SITE ADDRESS:	Wrexham Ind Est Wrexham LL13 9RD	INDICATOR TYPE:	Bilanciai EV2002
SERVICE REPORT NO:	22955	DATA PLATE:	
MANUFACTURER:	Bilanciai	INDICATOR SERIAL NO:	85604
TYPE:	TSE	MINIMUM CAPACITY (kg):	400
SIZE:	16M	MAXIMUM CAPACITY (kg):	50000
LOCATION:	Entrance	DIVISION (e) (kg):	20
		PRINTER SERIAL NO:	N/A
		TARE FACILITY:	Disabled

Accuracy of Zero, Linearity/Hysteresis, Discrimination & Comparison Tests = *

Approximate Test Interval (e)	MPE (e)	Actual Load (kg)	Indicator Up	Display Error (e)	True Error (e)	Indicator Down	Display Error (e)	True Error (e)	SL
ZERO	0.50	0	0	0.00		0	0.00		
2	0.50	40	40	0.00	0.00	44	0.20	0.20	
20	0.50	440	438	-0.10	-0.10	442	0.10	0.10	
500	1.0	10040	10038	-0.10	-0.10	10038	-0.10	-0.10	
1000	2.0	20040	20038	-0.10	-0.10	20038	-0.10	-0.10	SL1
1250	2.0	25040	25036	-0.20	-0.20	25036	-0.20	-0.20	
2000	2.0	40040	40036	-0.20	-0.20	40036	-0.20	-0.20	SL2
2200	3.0	44040	44034	-0.30	-0.30	44034	-0.30	-0.30	
SL1		20040							
SL2									
PASSED	Yes								

SL - Substitute Load

NOT TESTED AT MAX CAPACITY, BALLAST NOT PROVIDED

REPEATABILITY TEST (Zero Track On)
80%-MPE(e) 3.00

	Indicator	Indicator	Indicator
80%	44040	44040	44040
PASSED	Yes		

ECCENTRIC LOAD TEST - MPE (e): 1.0

Position	1	2	3	4	5	6	7	8	9	10
Test Load	8040	8040	8040	8040						
Indicator	8038	8038	8042	8042						
Error (e)	-0.10	-0.10	0.10	0.10						
PASSED	Yes									

LOADCELL DATA

Number	4
Make	Bilanciai
Type	CPR (Analogue)
Divisions	3000
Test cert	R60/2000-GB1-05.07
Conformity	Yes
PASSED	Yes

COMPARISON TEST

Printer	N/A
Remote	N/A
PC	N/A
Other	N/A
PASSED	N/A

CUSTOMER CONTACT: Matt Kirk

email: mattkirk@ashgrouppltd.co.uk

PHYSICAL CONDITION: Good
TEST WEIGHTS USED: PM1-28
DT1-17

AUTHORISED PERSON: Alan Dickson
CERTIFICATE NO: T286351
04239

CALIBRATION DATE: 28 November 2016

NEXT CALIBRATION DATE: 30 November 2017

SIGNATURE: *D. Campbell*